

**BLANK MATERIAL, AdLIGO COMPENSATION PLATE**

APPROVALS	DATE	REV	DCN NO.	BY	CHECK	DCC	DATE
AUTHOR: V. Parames	01/23/08	A	E080038-A				
CHECKED: G Billingsley	04/04/08	A					
APPROVED:							
DCC RELEASE							

**Applicable Documents**

LIGO-D080051-A AdLIGO CP Mirror Blank

MIL-G-174-B Glass, Optical

**Requirements**

Physical Dimensions	per LIGO- D080051-A AdLIGO CP Mirror Blank
Clear Aperture	Central 275 mm
Serial Number	Blanks shall be serialized as CPXX, where XX increments starting at 01
Material	Low Inclusion Fused Silica
Final shaping	Shaping shall be performed using a progression of grit size ending with a 320 or smaller grit tool.
Defect depth	Maximum on any surface or corner is less than 0.5 mm
Index Homogeneity in central 110 mm	$\leq 5 \times 10^{-7}$ P-V
Index Homogeneity in clear aperture	$\leq 2.5 \times 10^{-6}$ P-V
Birefringence	$\leq 1$ nm/cm within the clear aperture $\leq 5$ nm/cm outside the clear aperture
Bubble and Inclusion Cross section Within the clear aperture	Total $\leq 0.01 \text{ mm}^2/100\text{cm}^3$ Inclusions with a diameter of 10 $\mu\text{m}$ or less are not included in the total. Note all inclusions
Maximum inclusion diameter	$\leq 0.1$ mm
Striae	Class 1
OH Content	$\leq 1$ ppm

**BLANK MATERIAL, AdLIGO COMPENSATION PLATE**

Specification	Method	Frequency of Inspection	Data Delivered
Physical Dimensions	Visual Inspection	100%	Diameter Thickness
Serial number	Visual Inspection	100%	Inspection Report included with Certification
Material	Process Control Material Certification	100%	Inspection Report included with Certification
Defect depth	Visual Inspection	100%	Hand sketch indicating location and dimensions
Homogeneity	Interferometric Measurement	100%	Inspection Report included with Certification
Birefringence	MIL-G-174 Section 4.4.5	100%	Inspection Report included with Certification
Inclusions	Visual Inspection	100%	Hand sketch indicating location and dimensions
Striae	MIL-G-174 Section 4.4.6, method 1 or 2 (in optical axis only)	100%	Inspection Report included with Certification
OH content	Measurement	100%	Inspection Report included with Certification

**Table 1: MEASUREMENT MATRIX: FREQUENCY AND METHOD**