LIGO

LASER INTERFEROMETER GRAVITATIONAL WAVE OBSERVATORY

SPECIFICATION

E1000595 -V1 Drawing No Vers.

Sheet 1 of 2

aLIGO ISC Optics:

1" High Reflectors @ 1064nm

APPROVALS	DATE	RE V	DCN NO.	BY	CHECK	DCC	DATE
AUTHOR: L. BARSOTTI	10-14-10	•					
CHECKED:							
APPROVED: P.FRITSCHEL							
DCC RELEASE			`				

1 Description

1" Ø Flat/Flat high reflector @ 1064nm

2 Material

Corning HPFS 7980 (high purity fused silica, UV grade) Grade 0A (Low inclusion class: <0.3 mm² cross section, 0.1 mm max. size; Homogeneity < 1ppm)

3 Dimensions

1"Ø +.000/-.005" X .250" ± .020" tk., Plano / Plano

4 Surface Roughness

Side 1

Super polish

Surface Roughness: <1Å RMS in CA

Surface Quality: 10-5

Side 2

Commercial Polish

Surface Roughness: <5Å RMS in CA

Surface Quality: 20-10

5 Surface Figure

Side 1

Flat $< \lambda/10$ at 632.8 over central 80%

Side 2

Flat $< \lambda/10$ at 632.8 over central 80%

LASER INTERFEROMETER GRAVITATIONAL WAVE OBSERVATORY

SPECIFICATION

E1000595 -V1

Drawing No Vers.

Sheet 2 of 2

aLIGO ISC Optics:

1" High Reflectors @ 1064nm

6 Coating

Wavelength: 1064nm

Angle of incidence: 45°±5° (best effort for wider AOI range)

Side 1

 $R \ge 99.995\%$ @ 1064nm (best effort) for **s** and **p**-polarization

Side 2

AR coating, R < 1% @ 1064nm (best effort) for $\bf s$ and $\bf p$ -polarization

Serial numbers and registration marks shall be scribed or etched on the barrel of the optic for in-vacuum use

Coating vendor to provide:

- 1. Three spectrophotometer graphs of the reflectance and transmittance of the HR coatings; one covering the spectrum from 500nm to 1200nm; the others, with increased sensitivity, showing wavelengths from 900nm to 1100nm and from 500nm to 600nm
- 2. Spectrophotometer graphs of the reflectance of the AR coating taken as cited above.