


**CALIFORNIA INSTITUTE OF TECHNOLOGY**  
**Laser Interferometer Gravitational Wave Observatory (LIGO) Project**

To: A. Lazzarini  
From: L. Jones   
Phone: 2970  
Refer to: LIGO-E950116-00-B  
Date: 12/27/95

Subject: CBI's pump port positioning accuracy

CBI has assured me that the axial positioning tolerances for beam tube pump ports will be less than +/- 1". This item resolves all TBDs in the BT/CC document that were assigned to the beam tube, as I understand it.

aaa:bbb

cc:

F. Asiri

M. Coles

D. Coyne

O. Matherny

G. Stapfer

Chronological File

Document Control Center