

COC QA TEST REPORT

Sheet 1 of 3

ITM04 HR Integrated Scatter

Test Date	June 13-16, 2012		
Author(s)	Liyuan Zhang, Margot Phelps, Patrick Murphy, GariLynn Billingsley		
Approval(s)			
Specification Doc.	LIGO-E0900041	Specification	Not specified in E0900041
Procedure Doc.	LIGO-E1000863		
1 st Scan (Mean ± Error*)	11 ± 2 ppm		
2 nd Scan (Mean ± Error*)	10 ± 2 ppm		
Conclusion	Qualified.		

*An estimated systematic error from previous multiple measurements with one LIGO-1 mirror.

Discussions and Comments:

Two scans of TIS measurement are carried out on ITM04 HR surface with positioning the arrow on barrel at Y+ direction. One is a full coverage scan over a 48 mm diameter aperture with the beam and step sizes of 0.3 mm and other a sampling scan over a 160 mm diameter aperture with the same beam size of 0.3 mm and step size of 1.0 mm, as shown in Figs. 1 and 2 respectively. The system was checked and calibrated with a 1" HR mirror (No.8128).

It's interesting to note there are two low scatter line-like features at center as shown in Fig. 1, it seems they are correlated with the features shown in the surface finish data (C1000472).



LASER INTERFEROMETER GRAVITATIONAL WAVE OBSERVATORY

Document No

COC QA TEST REPORT

Sheet 2 of 3

Rev.

ITM04 HR Integrated Scatter





LASER INTERFEROMETER GRAVITATIONAL WAVE OBSERVATORY

Document No

COC QA TEST REPORT

Sheet 3 of 3

Rev.

ITM04 HR Integrated Scatter



LIGO Form F0900006-v2