




D1201237 Top clamp, Dual suspended seismometer platform, PART PDM REV: X-000, DRAWING PDM REV:

8 7 6 5 4 3 2 1

NOTES CONTINUED:
 ⑤ SCRIBE, ENGRAVE (A VIBRATORY TOOL MAY BE USED), LASER MARK OR MECHANICALLY STAMP (NO INKS OR DYES) DRAWING PART NUMBER, REVISION (AND VARIANT OR "TYPE" IF APPLICABLE) ON NOTED SURFACE OF PART FOLLOWED ON THE NEXT LINE WITH A THREE DIGIT SERIAL NUMBER. SERIAL NUMBERS START AT 001 FOR THE FIRST ARTICLE AND PROCEED CONSECUTIVELY. USE MINIMUM 0.12" HIGH CHARACTERS, UNLESS THE SIZE OF THE PART DICTATES SMALLER CHARACTERS. EXAMPLE: DXXXXXX-VY, TYPE-XX, S/N XXX

REV.	DATE	DCN #	DRAWING TREE #
v1	28 SEP 2012	E1200841	-
-	-	-	-
-	-	-	-



NOTES AND TOLERANCES: (UNLESS OTHERWISE SPECIFIED)				LIGO CALIFORNIA INSTITUTE OF TECHNOLOGY MASSACHUSETTS INSTITUTE OF TECHNOLOGY		PART NAME	
DIMENSIONS ARE IN INCHES TOLERANCES: .XX ± .015 .XXX ± .005 ANGULAR ± 0.5°				1. INTERPRET DRAWING PER ASME Y14.5-1994. 2. DO NOT REMOVE SHARP EDGES. 3. DO NOT SCALE FROM DRAWING. 4. APPROXIMATE WEIGHT = 6 LB.		Top clamp	
						SYSTEM ADVANCED LIGO SUB-SYSTEM SEI	
MATERIAL 6061-T6 Al		FINISH 63 μinch		NEXT ASSY D1201254 / D1201255		DESIGNER P. KNOEHE 13 SEP 2012	SIZE DWG. NO. B D1201237
				DRAFTER P. KNOEHE 13 SEP 2012		REVISION v1	
				CHECKER MATICHARD 28 SEP 2012		SCALE: 1:3	PROJECTION:
				APPROVAL MATICHARD 28 SEP 2012		SHEET 1 OF 1	

8 7 6 5 4 3 2 1