



W. L. Gore shielded 4-conductor cable to be held in place in the SRI MicroD connection with Epoxy Put EP-1730-1 UVB compatible epoxy. Shield must be secured inside the connector body with the epoxy.

Title				Last Edited:	
<b>Cable-DCPD Preamps to Breadboard Bracket</b>				LIGO Laboratory California Institute of Technology, Massachusetts Institute of Technology	
Size: A	DCC Number: D1300370	Revision: v1	Engineer:	Date: 7/11/2013	LIGO Date: 12:38:00 PM Sheet 1 of 1
File: C:\Users\cmh\Documents\drab_ligo_files\Chab\Altium\GMC_cable\41300370_SchDoc					