

LIGO Laboratory / LIGO Scientific Collaboration

LIGO-E1300587-v2

ADVANCED LIGO

19-Nov-13

HAUX H1-IM3 test report

Giacomo Ciani

Distribution of this document: LIGO Science Collaboration

This is an internal working note of the LIGO Project.

California Institute of Technology LIGO Project – MS 18-34 1200 E. California Blvd. Pasadena, CA 91125 Phone (626) 395-2129 Fax (626) 304-9834 E-mail: info@ligo.caltech.edu

LIGO Hanford Observatory P.O. Box 1970 Mail Stop S9-02 Richland WA 99352 Phone 509-372-8106 Fax 509-372-8137 Massachusetts Institute of Technology LIGO Project – NW17-161 175 Albany St Cambridge, MA 02139 Phone (617) 253-4824 Fax (617) 253-7014 E-mail: info@ligo.mit.edu

LIGO Livingston Observatory P.O. Box 940 Livingston, LA 70754 Phone 225-686-3100 Fax 225-686-7189

http://www.ligo.caltech.edu/

Contents

1	Int	Introduction				
	1.1	Sus	spension data			
	1.2	Ap	plicable Documents			
	1.2	2.1	LIGO Documents			
2	Su	mma	ry of tests			
3	Tests results					
	3.1	OS	EMs OLV			
	3.2	DC	Pointing			
3.3 OSEMs range and linearity			EMs range and linearity			
	3.3	8.1	Mirror rotation Vs Actuation			
	3.3	8.2	OSEMs readout Vs Displacement			
	3.4	Lin	ear spectra, no ECDs7			
	3.5	Me	asured resonances			
	3.6	Tra	nsfer functions, no ECDs			
	3.6	5.1	Length excitation			
	3.6	5.2	Pitch excitation			
	3.6	5.3	Yaw excitation			
	3.7	Lin	ear spectra, with ECDs 12			
	3.8	Tra	nsfer functions, with ECDs			
	3.8	8.1	Length excitation			
3.8.2		3.2	Pitch excitation			
	3.8	8.3	Yaw excitation			
	3.9	Qua	ality factors with ECDs			
	3.9	0.1	Bounce			
	3.9	0.2	Trans			
	3.9	9.3	Roll			
	3.10	S	tructural resonances			

1 Introduction

This document summarizes the results of tests conducted to verify L1 HAM Auxiliary suspensions' compliance with requirements, as well as other useful information.

1.1 Suspension data

IFO: H1 Suspension name: IM3 Suspension SN: 008 Installed optics: PMMT2-02 UL OSEM SN: 377 LL OSEM SN: 292 UR OSEM SN: 404 LR OSEM SN: 309

https://ics-redux.ligo-la.caltech.edu/JIRA/browse/ASSY-D1000120-008

1.2 Applicable Documents

1.2.1 LIGO Documents

LIGO-T1200469, "HAUX test procedure and acceptance criteria"

2 Summary of tests

The following table helps to quickly identify in which condition the results of the tests reported in this document refer to.

Gray cells represent the minimum required condition for final testing. "X" indicates the conditions of the test which results are reported in this document.

	Та	Table		Electronics		Cables		sure	Result
	Test	HAM	Test	Final	Test	Final	Air	Vac	
OSEMs OLV		Х		Х		Х	Х		Passed
DC pointing	Х			Х	Х		Х		Passed
OSEMs calibr.	Х			X	Х		Х		Passed
PSDs, no ECDs	Х			X	Х		Х		Passed
TFs, no ECDs	Х			X	Х		Х		Passed
PSDs, with ECDs		X		X		X	Х		Passed
TFs, with ECDs		X		Х		Х	Х		Passed
Q measurements									Pending
B&K Hammering		X					Х		Passed

3 Tests results

3.1 OSEMs OLV

These measurements have been performed on HAM table with final electronics and cables in date 29-Nov-2012.

	UL Channel	LL Channel	UR Channel	LR Channel
UL OSEM	24080			
LL OSEM		24298		
UR OSEM			25009	
LR OSEM				22990

Requirements (from <u>T1200469</u>, § 2.2.4):

• >25k optimal, >20K acceptable

3.2 DC Pointing

This has been measured and corrected chamber-side in date 05-Sep-2012. See LHO aLog 4081.

Measured value: 0 ± 0.5 mrad

Requirements (from <u>T1200469</u>, § 2.1.1):

• Nominal value $(0) \pm 1$ mrad

3.3 OSEMs range and linearity

These measurements have been taken chamber-side in date 04-Sep-2012. See LHO aLog 4054.

Please note: these measurements have been taken with the HAUX coil driver (D1100117-v2 HAM-A coil drivers) "as designed". During commissioning, proposals have been made to reduce the electronics output gain such that the actuation range of the HAUX would better correspond to what is actually needed (see for example LLO aLog 5213). However, as of the time of writing this document, modifications have not yet been definitively approved or applied to all coil driver boxes and are not considered part of these acceptance tests.

<u>Passed</u>

Passed



3.3.1 Mirror rotation Vs Actuation

Requirements (from <u>T1200469</u>, § 2.1.1):

• > 10 mrad for full actuation (32000 counts)

Passed

3.3.2 OSEMs readout Vs Displacement



There are no requirements associated with this measurement. It is rather intended to be a sanity check looking for strange (e.g. non linear) behaviors and differences between OSEMs. *Passed*

3.4 Linear spectra, no ECDs

These measurements have been taken chamber-side in date 31-Aug-2012. See aLog 4034.

Data is stored in the SUS SVN repository:

HAUX\H1\Common\1030461878_IMall_PSD_2mHz_ECDno_DAMPno_Barrel.xml



There are no requirements associated with this measurement. It is rather intended to be a sanity check looking for strange behaviors and differences between OSEMs. As a reference, high frequency electronic noise is expect to be $\sim 10^{-4} \,\mu m/Hz^{1/2}$.

3.5 Measured resonances

These have been measured chamber-side in date 04-Sep-2012. See LHO aLog 4054.

 Yaw:
 0.740 Hz

 Pitch/Length 1:
 0.910 Hz

 Pitch/Length 2:
 1.021 Hz

 Transverse:
 0.996 Hz

 Bounce:
 6.14 Hz

 Roll:
 9.03 Hz

Requirements (from <u>T1200469</u>, § 2.1.1):

- Pitch, Yaw, Length < 10 Hz (mandatory)
- Transverse, Bounce, Roll < 10 Hz (recommended)

<u>Passed</u> Passed

3.6 Transfer functions, no ECDs

These have been measured chamber side in date 31-Aug-2012. See LHO aLog 4034.

Data is stored in the SUS SVN repository:

HAUX\H1\Common\1030842823_IMall_TF-L_1e5_5mHz_ECD2_DAMPno_OSEMoffset_Barrel.txt HAUX\H1\Common\1030844369_IMall_TF-P_5e3_5mHz_ECD2_DAMPno_OSEMoffset_Barrel.txt HAUX\H1\Common\1030845884_IMall_TF-Y_5e3_5mHz_ECD2_DAMPno_OSEMoffset_Barrel.txt

Please note:

- Proper diagonalization of AOSEMs actuation and readout had not been performed at this stage; thus, cross-coupling between different DoF is visible.
- The "model" curve represents the TF obtained from the Mathematica model using nominal values for all parameters.
- Calibration of the measured data is done using nominal values for all elements involved in the actuation/readout chain.
- Due to the small weight of the HAUX optics and the need to perform testing in a clean environment under flowing filtered air, many TFs are affected by a comparatively high level of noise.

There is no quantitative requirement associated with this measurement. TFs are expected to be consistent with the model (see, <u>T1200469</u>, § 2.1.5), although close matching of resonances is not necessarily expected.

In general, all resonances appear to be lower than predicted by the model. This is common to all HAUX instances and does not pose a problem from a performance point of view. The TFs are considered acceptable as long as they do not show abnormal behaviors that can suggest rubbing or similar problems.

3.6.1 Length excitation



No abnormal behavior observed.



3.6.2 Pitch excitation



No abnormal behavior observed.

<u>Passed</u>

3.6.3 Yaw excitation

No abnormal behavior observed.

Passed

3.7 Linear spectra, with ECDs

These measurements have been taken on the HAM table, with purge air on, in date 06-Dec-2012.

Data is stored in the SUS SVN repository:

HAUX\H1\Common\1038891763_IMall_PSD_2mHz_ECD2_DAMPno_AlignOffset_PurgeAirOn.xml

These measurements need to be repeated in vacuum when the occasion arises.

There are no requirements associated with this measurement. It is rather intended to be a sanity check looking for strange behaviors and differences between OSEMs. As a reference, high frequency electronic noise is expect to be $\sim 10^{-4} \,\mu m/Hz^{1/2}$.

3.8 Transfer functions, with ECDs

These measurements have been taken on the HAM table, with purge air on, in date 06-Dec-2012.

Data is stored in the SUS SVN repository:

HAUX\H1\Common\1038897341_IMall_TF-L_1e5_5mHz_ECD2_DAMPno_AlignOffset_PurgeAirOn.xml HAUX\H1\Common\1038901324_IMall_TF-P_5e3_5mHz_ECD2_DAMPno_AlignOffset_PurgeAirOn.xml HAUX\H1\Common\1038899227_IMall_TF-Y_5e3_5mHz_ECD2_DAMPno_AlignOffset_PurgeAirOn.xml

These measurements need to be repeated in vacuum when the occasion arises.

Please note:

- Proper diagonalization of AOSEMs actuation and readout had not been performed at this stage; thus, cross-coupling between different DoF can be visible.
- The "reference" curve represents the TF measured with no ECDs; it is the same plotted in section 3.6.
- Due to the small weight of the HAUX optics and the need to perform testing in a clean environment under flowing filtered air, many TFs are affected by a comparatively high level of noise.
- In principle, we are not interested in any passive damping of yaw, pitch and length, as they can be controlled actively. However, coupling with these DoFs is a known issue of the ECD system designed to damp the other DoFs.

There is no quantitative requirement associated with this measurement, which is mostly intended as a sanity check.

3.8.1 Length excitation

No abnormal behavior observed.

3.8.2 Pitch excitation

No abnormal behavior observed.

Passed

3.8.3 Yaw excitation

No abnormal behavior observed.

<u>Passed</u>

3.9 Quality factors with ECDs

Data for these measurements have been taken with different techniques and yielded mixed results. They need to be measured again in a more controlled and uniform way. This can be easily repeated without physically accessing the suspensions, but requires waiting for the right window of opportunity while the IFO is being commissioned.

3.9.1 Bounce

- 3.9.2 Trans
- 3.9.3 Roll

3.10 Structural resonances

Measurements have been taken on the HAM table, in final clamping configuration, in date 05-Mar-2013. See LHO a Log 5652.

10²

10¹

Requirements (from T1200469, § 2.1.4):

• All resonances >150 Hz

10⁰ Magnitude [(m / s²) / N] 10 10 10 10 H1SUSIM1 (HAUX), VAs OFF, Y Resp. to X Exc. Test3 H1SUSIM2 (HAUX), VAs OFF, Y Resp. to X Exc. Test3 H1SUSIM3 (HAUX), VAs OFF, Y Resp. to X Exc. Test3 H1SUSIM4 (HAUX), VAs OFF, Y Resp. to X Exc. Test3 10 10^{⊸∟} 50 70 80 90 100 60 200 300 400 Frequency [Hz] B&K Hammer Measurement Comparison 10² 10¹ 10 Magnitude [(m / s²) / N] 10 H1SUSIM1 (HAUX), VAs OFF, Y R 10 HISUSIMI (HAUX), VAS OFF, Y Resp. to Y Exc. Testa HISUSIM3 (HAUX), VAS OFF, Y Resp. to Y Exc. Testa HISUSIM3 (HAUX), VAS OFF, Y Resp. to Y Exc. Testa HISUSIM4 (HAUX), VAS OFF, Y Resp. to Y Exc. Testa 70 80 90 100 60 300 400 200 Frequency [Hz]

B&K Hammer Measurement Comparison

<u>Passed</u>