




D2000453 Peek Feedthrough x2  
Band-it Microband Glenair# 601-025 x2



AccuGlass Wire Type 112676 or Equivalent  
Wire Gauge: 28 AWG  
Wire Resistance: 0.244 ohms/meter  
Wire Type: UHV Kapton Insulated  
PEEK Overbraid Type 111167 or Equivalent



Strain Relief Accu-Glass  
Part Number 100465  
A+ Version Assembly (two posts only)



Accu-Glass  
Part Number 100421



Last Edited: 5/17/2023

Title <b>OMCB D25M - to - QPDx2</b>		LIGO Laboratory California Institute of Technology Massachusetts Institute of Technology		LIGO
Size: A	DCC Number: D2300123	Revision: v2	Engineer: Dean Schaeztl	
File: C:\Dean\M2100130 OMC FDR Update\Altium Files\D2300123-v2 OMCB D25M - to - QPDx2.SchDoc				
Sheet 14 of 17				